

Chinese Hamster Ovary Host Cell Proteins, CHO HCP ELISA Kit

96 Tests

Operating instruction

FOR RESEARCH USE ONLY; NOT FOR THERAPEUTIC OR DIAGNOSTIC APPLICATIONS! PLEASE READ THROUGH ENTIRE PROCEDURE BEFORE BEGINNING!

Synonyms

CHO HCP, CHO-HCP, Hamster Ovary Host Cell Proteins, HCP

Search name

CHO HCP ELISA KIT, CHO-HCP ELISA KIT, Chinese Hamster Ovary Host Cell Proteins ELISA KIT, HCP ELISA KIT,

Intended use

This immunoassay kit allows for the in vitro quantitative determination of Chinese Hamster Ovary (CHO) Host Cell Proteins concentrations in products manufactured by recombinant expression in CHO host cells.

Test principle

The microtiter plate provided in this kit has been pre-coated with an antibody specific to CHO Host Cell Proteins. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody preparation specific for CHO Host Cell Proteins and Streptavidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. Then a TMB substrate solution is added to each well. Only those wells that contain CHO Host Cell Proteins, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The concentration of CHO Host Cell Proteins in the samples is then determined by comparing the O.D. of the samples to the standard curve.

Materials and components

Reagent	Quantity
Assay plate	1
Standard	600 μ L
Reporting antibody (1mg/mL)	150 μ L
Streptavidin-HRP Conjugate (4 μ g/mL)	375 μ L
Dilution Buffer Composition	0.4g
10 \times PBS-T	45mL
TMB Substrate	15mL
Stop Solution	15mL
Plate sealer for 96 wells	1
Instruction	1

Other supplies required

Microplate reader.

Pipettes and pipette tips.

EP tube

Deionized or distilled water.

Storage of the kits

The **Assay Plate, Standard, Reporting antibody and Streptavidin-HRP Conjugate** should be stored at -20 $^{\circ}$ C upon being received.

After receiving the kit, **Substrate should be always stored at 4 $^{\circ}$ C**. Other reagents are kept according to the labels on vials. But for long term storage, please keep the whole kit at -20 $^{\circ}$ C (Except the substrate). The unused strips should be kept in a sealed bag with the desiccant provided to minimize exposure to damp air. The test kit may be used throughout the expiration date of the kit (six months from the date of manufacture). Opened test kits will remain stable until the expiring date shown, provided it is stored as prescribed above.

Note: Because of the inherent stability of lyophilized material, Nova lifetech may ship these materials at ambient temperature.

Limitations of the procedure

1. Nova lifetech. is only responsible for the kit itself, but not for the samples consumed during the assay. The user should calculate the possible amount of the samples used in the whole test. Please reserve sufficient samples in advance.
2. The kit should not be used beyond the expiration date on the kit label.
3. Do not mix or substitute reagents with those from other lots or sources.
4. If samples generate values higher than the highest standard, further dilute the samples with the Dilution Buffer and repeat the assay. Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.

Reagent preparation

10×PBS-T - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 45 mL of Wash Buffer Concentrate into deionized or distilled water to prepare 450 mL of 1×PBS-T. **Dilution Buffer** – Add all the Dilution Buffer Composition (0.4g) to 40mL of 1×PBS-T, and mix uniformly.

Standard – The concentration of Standard stock is 2430 ng/mL. Suck out 300µL of the stock into a clean eppendorf tube, and then add 600µL of Dilution Buffer to produce a standard solution of 810 ng/mL (serves as the high standard). Allow the standard solution to sit for a minimum of 15 minutes with gentle agitation prior to making serial dilutions (270ng/mL, 90ng/mL, 30ng/mL, 10ng/mL, 3.3ng/mL) at the dilution of 1:3. The Dilution Buffer serves as the zero standard (0 ng/mL).

Reporting antibody working solution- Dilute to the working concentration (10µg/mL) using Dilution Buffer at the dilution of 1:100.

Streptavidin-HRP Conjugate working solution- Dilute to the working concentration (0.1µg/mL) using Dilution Buffer at the dilution of 1:40.

Assay procedure

Allow all reagents to reach room temperature (Please do not dissolve the reagents at 37°C directly.). **All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming.** Keep appropriate numbers of strips for 1 experiment and remove extra strips from microtiter plate. Removed strips should be resealed and stored at 4°C until the kits expiry date. Prepare all reagents, working standards and samples as directed in the previous sections. Please predict the concentration before assaying. If values for these are not within the range of the standard curve, users must determine the optimal sample dilutions for their particular experiments.

1. Add 100 µl of **Standard**, Blank, or Sample per well. Cover with the Plate sealer. Incubate for 1.5 hours at room temperature.
2. Remove the liquid of each well. Aspirate each well and wash, repeating the process three times for a total of three washes. Wash by filling each well with 1×PBS-T (approximately 250 µl) using a squirt bottle, multi-channel pipette, manifold dispenser or autowasher. and let it sit for 1~2 minutes. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining 1×PBS-T by aspirating or decanting. Invert the plate and blot it against clean paper towels.
3. Add 100 µl of **Reporting antibody working solution** to each well. Cover with the Plate sealer. Incubate for 45min at room temperature.
4. Repeat the aspiration/wash process for 5 times as conducted in step 2
5. Add 100 µl of **Streptavidin-HRP Conjugate working solution** to each well. Cover with a new Plate sealer. Incubate for 30min at room temperature.
6. Repeat the aspiration/wash process for 5 times as conducted in step 2.
7. Add 100 µl of **TMB Substrate** to each well. Cover with a new Plate sealer. Incubate about **15 minutes**. Protect from light.
7. Add 100 µl of **Stop Solution** to each well. If color change does not appear uniform, gently tap the plate to ensure thorough mixing.
8. Determine the optical density of each well at once, using a microplate reader set to 450 nm.

Note:

1. Absorbance is a function of the incubation time. Therefore, prior to starting the assay it is recommended that all reagents should be freshly prepared prior to use and all required strip-wells are secured in the microtiter frame. This will ensure equal elapsed time for each pipetting step, without interruption.
2. Please carefully reconstitute Standards or working Reporting antibody and Streptavidin-HRP Conjugate according to the instruction, and avoid foaming and mix gently until the crystals have completely dissolved. **The reconstituted Standards, Reporting antibody and Streptavidin-HRP Conjugate can be used only once.** This assay requires pipetting of small volumes. To minimize imprecision caused by pipetting, ensure that pipettors are calibrated. It is recommended to suck more than 10µl for once pipetting.
3. To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary. Do not allow wells to sit uncovered for extended periods between incubation steps. Once reagents have been added to the well strips, DO NOT let the strips DRY at any time during the assay.
4. For each step in the procedure, total dispensing time for addition of reagents to the assay plate should not exceed 10 minutes.
5. To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
6. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbance readings.
7. Duplication of all standards and specimens, although not required, is recommended.
8. Substrate Solution is easily contaminated. Please protect it from light.
9. The web version of manual is only for reference, subject to the instruction shipping with the kit.

Specificity

This assay recognizes recombinant and natural CHO Host Cell Proteins. No significant cross-reactivity or interference was observed.

Note:

Limited by current skills and knowledge, it is impossible for us to complete the cross- reactivity detection between CHO Host Cell Proteins and all the analogues, therefore, cross reaction may still exist.

Detection Range

3.3 ng/mL -270 ng/mL.

Calculation of results

Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the x-axis against the concentration on the y-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the CHO Host Cell Proteins concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. It is recommended to use some related software to do this calculation, such as curve expert 1.3. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

Important note:

1. Limited by the current condition and scientific technology, we can't completely conduct the comprehensive identification and analysis on the raw material provided by suppliers. So there might be some qualitative and technical risks to use the kit
2. The final experimental results will be closely related to validity of the products, operation skills of the end users and the experimental environments. Please make sure that sufficient samples are available.
3. Kits from different batches may be a little different in detection range, sensitivity and color developing time. Please perform the experiment exactly according to the instruction attached in kit while electronic ones from our website is only for information.
4. There may be some foggy substance in the wells when the plate is opened at the first time. It will not have any effect on the final assay results.
5. Do not remove microtiter plate from the storage bag until needed.
6. A microtiter plate reader with a bandwidth of 10nm or less and an optical density range of 0-3 OD or greater at 450nm wavelength is acceptable for use in absorbance measurement.
7. Use fresh disposable pipette tips for each transfer to avoid contamination.
8. Do not substitute reagents from one kit lot to another. Use only the reagents supplied by manufacturer.
9. Even the same operator might get different results in two separate experiments. In order to get better reproducible results, the operation of every step in the assay should be controlled. Furthermore, a preliminary experiment before assay for each batch is recommended.
10. Each kit has been strictly passed Q.C test. However, results from end users might be inconsistent with our in-house data due to some unexpected transportation conditions or different lab equipments. Intra-assay variance among kits from different batches might arise from above factors, too.
11. Kits from different manufacturers for the same item might produce different results, since we haven't compared our products with other manufacturers.
12. The instruction manual also suit for the kit of 48T, but all reagents of 48T kit is reduced by half.
13. Valid period: six months.

Precaution

The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material.